

Notice of Allowability

Application No.

10/679,043

Applicant(s)

FOOS ET AL.

Examiner

Josiah Cocks

Art Unit

3749

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to papers filed through August 27, 2007.
2. ☒ The allowed claim(s) is/are 14-19, 31, 41, 44-48, 55-66, 68 and 70-72.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some* c) ☐ None of the:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).


* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
- (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
- 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
- (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statements (PTO/SB/08), Paper No./Mail Date _____
- ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material
- ☐ Notice of Informal Patent Application
- ☐ Interview Summary (PTO-413), Paper No./Mail Date _____
- ☒ Examiner's Amendment/Comment
- ☐ Examiner's Statement of Reasons for Allowance
- ☒ Other attached PTOL-413B


Josiah Cocks
Primary Examiner
Art Unit 3749

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Tait Swanson on September 4, 2007. See the attached Examiner-Initiated Interview Summary form PTOL-413B for a description of the reasons for examiner's amendment.

2. The application has been amended as follows:

IN THE CLAIMS:

In **claim 61**, line 9, after the term "orientation" deleted the "." and inserted:

--; wherein the valve passage is coupled to a first fluid inlet, a second fluid inlet, and a fluid outlet, wherein the valve receives fluid from the first fluid inlet when in the first orientation and from the second fluid inlet when in the second orientation, and the valve outputs the fluid to the fluid outlet in both the first and second orientations.--

Claim 67 has been cancelled.

DETAILED ACTION

Response to Amendment

3. Receipt of applicant's amendment filed August 27, 2007 is acknowledged.

Allowable Subject Matter

4. In light of the amendment filed August 27, 2007 and the examiner's amendment attached to this Office action, claims 14-19, 31, 41, 44-48, 55-66, 68, and 70-72 are allowed.

USPTO Contact Information

5. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Josiah Cocks whose telephone number is (571) 272-4874. The examiner can normally be reached on M-F 8:00-5:30.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Steven McAllister, can be reached (571) 272-6785. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would

Art Unit: 3749

like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

jcc
September 4, 2007



JOSIAH COCKS
PRIMARY EXAMINER
ART UNIT 3749